

SenTix[®] 950(-P)

SenTix[®] 980(-P)

SenTix[®] Micro 900(-P)



SenTix[®] 950 / 980 / Micro 900(-P)

pH ELECTRODE WITH LIQUID ELECTROLYTE



a xylem brand

Copyright © 2017, Xylem Analytics Germany GmbH
Printed in Germany.

1 General information

Automatic sensor recognition

The sensor electronics with the stored sensor data are in the connecting head of the electrode. The data include, among other things, the sensor type and series number. In addition, the calibration data are stored in the sensor with each calibration and the calibration history is recorded (the last 10 calibrations). The data is recalled by the meter when the sensor is connected and is used for measurement and for measured value documentation.


Storing the calibration data in the sensor ensures that the correct slope and asymmetry are automatically used if the sensor is operated with different meters. On the other hand, different calibrated sensors can be used with one meter without the need to recalibrate.

The digital transmission technique guarantees the failure-free communication with the meter even with long connection cables. If the sensor firmware is enhanced by WTW, it can be updated via the meter.

2 Commissioning, measuring, calibration

2.1 Opening and closing the IDS plug connection

This section only applies to IDS plug variants (SenTix® ... -P).



Opening the plug connection

- If necessary, clean the plug connection.
- Open the locking device (step 1).
- Use your thumb and index finger to press the clips of the connector together, and pull the connector out of the plug (step 2).

Closing the plug connection

- Make sure that the plug connection is completely dry and clean.
- Align the guidance of the connector with the groove in the plug and insert the connector in the unlocked plug until it catches (step 3).
- Close the locking device (step 4).

2.2 Commissioning

Scope of delivery


- Electrode SenTix® 950 / 980 / Micro 900(-P)
- Operating manual

Commissioning

The electrode is filled with reference electrolyte solution in the factory. Prepare the electrode for measuring as follows:

- Open the refilling opening for the reference electrolyte solution. Depending on the model, the stopper of the refilling opening is an elastomer stopper or a slider.

The refilling opening must always be open during calibration and measurement!



- Remove the watering cap from the electrode tip. Possible salt deposits in the area of the watering cap do not affect the measuring characteristics and can easily be removed with deionized water.



Please keep the watering cap. It is required for the electrode to be stored. Always keep the watering cap clean.

- SenTix® 950(-P): Remove any gas bubbles behind the pH membrane by shaking.
- Connect the electrode to the meter.

SenTix® 950 / 980 /
Micro 900

– via the sensor cable to a free IDS connector on the meter

SenTix® 950 / 980 /
Micro 900-P

– via a connecting cable (accessory) to a free IDS connector on the meter

or

– wireless via an IDS WLM-S adapter (accessory) to a WLM-capable meter


Accessories for the connection of the SenTix® 980-P sensor to the meter: See chapter 7 WEAR PARTS AND ACCESSORIES.

Opening and closing the IDS plug connection, see section 2.1 OPENING AND CLOSING THE IDS PLUG CONNECTION.

- Calibrate the electrode according to the operating manual of the meter and observe the following rules while doing so:

2.3 Calibration and measurement: General rules

- Make sure the refilling opening for the reference electrolyte solution is open.
- Avoid the carryover of any solution (sample or buffer solution) from one measurement to the next by taking the following measures:
 - Shortly rinse the calibration and sample beakers with the solution the beakers are to be filled with next.
 - Between measurements, rinse the electrode with the solution that follows. Alternatively, you can also rinse the electrode with deionized water and then carefully dab it dry.
- Immerse the electrode in the solution in a vertical or slightly tilted position. Make sure the immersion depth is correct. The junction must be completely submerged in the solution. The junction is in the area of the bottom end of the shaft (see figure). At the same time, the level of the reference electrolyte in the electrode must be at least 2 cm above the level of the solution.



- Provide approximately the same stirring conditions for measuring as for calibrating.



Prevent contact of the pH membrane to the beaker bottom to avoid scratches on the pH membrane.

Subsequent calibrations

The frequency of subsequent calibrations depends on the application. The meters provide an option where you can enter a calibration interval. After the calibration interval has expired, the meter will automatically remind you of the due calibration.

3 Storage

During short measuring breaks

With the refilling opening open, immerse the electrode in reference electrolyte (KCl 3 mol/l, Ag⁺ free). Prior to the next measurement, shortly rinse the electrode with the test sample or deionized water.



Prevent contact of the pH membrane to the beaker bottom to avoid scratches on the pH membrane.

Overnight or longer

Put the clean electrode in the watering cap that is filled with reference electrolyte (KCl 3 mol/l, Ag⁺ free) and close the refilling opening.

NOTE

pH electrodes must not be stored dry or in deionized water. The electrode could be permanently damaged by this. If the liquid in the watering cap has dried up, condition the electrode in reference electrolyte (KCl 3 mol/l, Ag⁺ free) for at least 24 hours.



During longer storing periods, salt sediments may develop on the watering cap. They do not affect the measuring characteristics and can easily be removed with deionized water when the electrode is put into operation again.

4 Aging

pH electrodes are consumables. Every pH electrode undergoes a natural aging process. With aging, the responding behavior becomes slower and the electrode slope and asymmetry change. Moreover, extreme operating conditions can considerably shorten the lifetime of the electrode. These are:

- Strong acids or lyes, hydrofluoric acid, organic solvents, oils, fats, bromides, sulfides, iodides, proteins
- High temperatures
- High changes in pH and temperature.


The warranty does not cover failure caused by measuring conditions and mechanical damage.

5 Maintenance and cleaning

Refilling the reference electrolyte

During operation, a small amount of reference electrolyte leaks through the junction from the electrode into the test sample. If the level of reference electrolyte becomes too low with time, refill it through the refilling opening. Refilling is very easy with the enclosed dropping bottle. Proceed as follows:

- Cut off the tip of the dropping bottle at a right angle until the opening in the tip can be seen
- Open the refilling opening of the electrode
- Press the tip of the dropping bottle into the refilling opening while turning it slightly
- Pump the reference electrolyte in the shaft using the dropping bottle
- Pull the dropping bottle out of the refilling opening while turning it slightly as necessary.



Cleaning

Remove water-soluble contamination by rinsing with deionized water. Remove other contamination as follows:

Contamination	Cleaning procedure
Fat and oil	Rinse with water containing household washing-up liquid
Lime and hydroxide deposits	Rinse with citric acid (10 % by weight)
Proteins	Immerse in pepsin cleaning solution PEP/pH for approx. 1 hour. <u>Note:</u> Make sure the level of the reference electrolyte is above that of the cleaning solution.

NOTE

Hydrofluoric acid, hot phosphoric acid and strong alkaline solutions destroy the glass membrane.

After cleaning Rinse the electrode with deionized water and condition it in reference electrolyte solution for at least 1 hour. Then recalibrate the electrode.

6 Technical data

Measurement	pH measuring range	0.000 ... 14.000
	Allowed temperature range (°C)	0 ... 80 °C (SenTix® 950(-P)) 0 ... 100 °C (SenTix® 980 / Micro 900(-P))
	Allowed temperature range (°F)	32 ... 176 °F (SenTix® 950(-P)) 32 ... 212 °F (SenTix® 980 / Micro 900(-P))
Accuracy of the IDS measuring technique	Measured parameter	Accuracy (± 1 digit)
	pH	± 0.004
	U [mV]	± 0.2
	T [°C]	± 0.1
General features	Reference electrolyte	KCl 3 mol/l, Ag ⁺ free
	Junction	Ceramic (SenTix® 950(-P)) Platinum (SenTix® 980 / Micro 900(-P))
	Temperature sensor	Integrated NTC 30 (30 kΩ at 25 °C / 77 °F)
Connection cable	Lengths	SenTix® 950 / 980 / Micro 900: 1,5 SenTix® 950 / 980 / Micro 900-P: 1,5 / 3 / 6 / 10 / 15 / 25 / 40 / 60 / 100 m
	Diameter	4.3 mm
	Smallest allowed bend radius	Fixed installation: 20 mm Flexible use: 60 mm
	Plug type	Socket, 4 pins
Shaft dimensions, material	Shaft length	120 mm (SenTix® 950 / 980(-P)) 65/130 mm (SenTix® Micro 900(-P))
	Shaft diameter	12 mm (SenTix® 950 / 980(-P)) 12/5 mm (SenTix® Micro 900(-P))
	Shaft material	Polyamide (SenTix® 950(-P)) Glass (SenTix® 980(-P), SenTix® Micro 900(-P))

IDS plug		<ul style="list-style-type: none"> ● Synthetic materials: Glass fiber reinforced Noryl, TPU, TPC-ET, POM, PVC, PEEK, PBT ● O-ring: FPM ● Contacts gold-Plated
IDS plug	Connection type	4-Pole, watertight plug connection with lock, reverse polarity protected
Storage	With watering cap; filled with KCl 3 mol/L, Ag ⁺ free	

7 Wear parts and accessories

Maintenance equipment

Description	Model	Order no.
Reference electrolyte solution 250 ml to fill the watering cap (KCl 3 mol/l, Ag ⁺ -free)	KCl-250	109 705
Pepsin cleaning solution 3 x 250 ml	PEP/pH	109 648

Connection cable SenTix® 950 / 980 / Micro 900(-P) - meter

Description	Model	Order no.
IDS connection cable, 1.5 m	AS/IDS-1.5	903 850
IDS connection cable, 3 m	AS/IDS-3	903 851
IDS connection cable, 6 m	AS/IDS-6	903 852
IDS connection cable, 10 m	AS/IDS-10	903 853
IDS connection cable, 15 m	AS/IDS-15	903 854
IDS connection cable, 20 m	AS/IDS-20	903 855
IDS connection cable, 25 m	AS/IDS-25	903 856
IDS connection cable, 40 m	AS/IDS-40	903 857
IDS connection cable, 60 m	AS/IDS-60	903 858
IDS connection cable, 100 m	AS/IDS-100	903 859
Wireless module for plug head sensor	IDS WLM-S	108 141

Radio connection SenTix® 950 / 980 / Micro 900(-P) - meter

Description	Model	Order no.
WLM capable IDS meter + radio module for IDS meter	see Internet	
Radio module for plug head sensor	IDS WLM-S	108 141

General accessories	Description	Model	Order no.
	Plastic arming for SenTix® 950 / 980(-P) pH electrodes	A pHLab/K	903 841

8 Disposal

At the end of its operational lifetime, the electrode (electronic waste) must be returned to the disposal or return system statutory in your country (electronic waste). If you have any questions, please contact your supplier.

What can Xylem do for you?

We're a global team unified in a common purpose: creating innovative solutions to meet our world's water needs. Developing new technologies that will improve the way water is used, conserved, and re-used in the future is central to our work. We move, treat, analyze, and return water to the environment, and we help people use water efficiently, in their homes, buildings, factories and farms. In more than 150 countries, we have strong, long-standing relationships with customers who know us for our powerful combination of leading product brands and applications expertise, backed by a legacy of innovation.

For more information on how Xylem can help you, go to xylem.com.



Service address:

Xylem Analytics Germany
Sales GmbH & Co. KG
WTW
Dr.-Karl-Slevogt-Str. 1
82362 Weilheim
Germany

Tel.: +49 881 183-325
Fax: +49 881 183-414
E-Mail wtw.rma@xylem.com
Internet: www.WTW.com



Xylem Analytics Germany GmbH
Dr.-Karl-Slevogt-Str. 1
82362 Weilheim
Germany